

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS FOLATE

Code 86402

Lot No. 016

2025-03

2024-03

REAGENTS

Label	Description	Lot.	Exp	No
D4560	FOLATE	016	2025-03	6X6
B4560	Calibratore FOLATE	016	2026-07	1
C4560	Control FOLATE	016	2026-07	1
G4560	FOLATE Sample Diluent/Negative Control	016	2026-07	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	350	> 100
Control (ng/ml)	3,1	2,0-5,4
FOLATE Sample Diluent/Negative Control (ng/ml)	<2,0	< 2,0


CALIBRATOR

9	9	1	0	1	6	6	8	1	3	2	0	4	3	2	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature




CONTROL SERUM

9	9	1	0	1	6	6	1	9	2	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 19.03.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.