



QUALITY CONTROL CERTIFICATE

KIT NAME SS-B - CHORUS

Code	86024			
Lot No.	016	2025-03	2024-03	

REAGENTS

Label	Description	Lot	Exp.	NO
D3201	SS-B	016	2025-03	6X6
B3201	CALIBRATORE SS-B	016	2025-03	1
C3201	CONTROL SERUM SS-B	016	2025-03	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	448	≥ 100
Control Serum (AU/mL)	32,9	25,0-60,0

CALIBRATOR

1	9	4	0	1	6	6	3	2	2	3	4	3	8	5	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature MB

194016632234365100  
SS-B  
CALIBRATOR 0,175 mL IVD  
LOT 016  
2025-03  
DIESSE Monteriggioni Italy CE

CONTROL SERUM

1	9	4	0	1	6	6	1	8	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature MB

194016618340000001  
SS-B  
CONTROL SERUM 0,425 mL IVD  
LOT 016  
2025-03  
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 21/03/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.