

DIESE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME a-TPO - CHORUS

Code 86074/12

Lot No. 003

2025-01

2024-01

## REAGENTS

| Label | Description         | Lot. | Exp.    | NO  |
|-------|---------------------|------|---------|-----|
| D3591 | a-TPO               | 003  | 2025-01 | 2x6 |
| B3591 | CALIBRATORE a-TPO   | 003  | 2025-01 | 1   |
| C3591 | CONTROL SERUM a-TPO | 003  | 2025-01 | 1   |

## CONTROLS

|  | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

|                       |       |             |
|-----------------------|-------|-------------|
| Calibrator (mOD)      | 331   | $\geq 100$  |
| Control Serum (IU/mL) | 391,2 | 237,5-781,3 |

## CALIBRATOR


|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 3 | 0 | 1 | 0 | 0 | 3 | 6 | 6 | 2 | 5 | 4 | 1 | 2 | 6 | 5 | 7 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES



Operator Signature *FB*



## CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 3 | 0 | 1 | 0 | 0 | 3 | 6 | 2 | 7 | 5 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES



Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *FB*

DATE 21/03/2024

## WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.