

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19-sc, 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME snRNP-C - CHORUS

Code 86092

Lot No. 016

2025-03

2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D7441	snRNP-C	016	2025-03	6X6
B7441	CALIBRATORE snRNP-C	016	2025-03	1
C7441	CONTROL SERUM snRNP-C	016	2025-03	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	430	≥ 100
Control Serum (AU/mL)	62,8	30,0-89,1

CALIBRATOR

1	9	1	0	1	6	6	1	1	8	2	6	3	8	3	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature *HB*



CONTROL SERUM

1	9	1	0	1	6	6	1	8	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *HB*



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *[Signature]*

DATE 26/03/2026

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.