

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME dsDNA-G - CHORUS

Code 86032

Lot No. 012

2025-02

2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D3281	dsDNA-G	012	2025-02	6X6
B3281	CALIBRATORE dsDNA-G	012	2025-02	1
C3281	CONTROL SERUM dsDNA-G	012	2025-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	685	≥ 100
Control Serum (IU/mL)	43,4	34,4-79,2

CALIBRATOR

2 5 1 0 1 2 6 4 1 8 3 0 5 7 3 3 0 0

Label checked with barcode reader

YES ☒

Operator Signature MB



CONTROL SERUM

2 5 1 0 1 2 6 1 3 2 5 0 0 0 0 0 0 1

Label checked with barcode reader

YES ☒

Operator Signature MB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... MB

DATE 22.03.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.