

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME HERPES SIMPLEX 1+2 IgM -CHORUS

Code 81021

Lot No. 010

2025-02

2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D2221	Herpes simplex 1+2 IgM	010	2025-02	6x6
B2221	CALIBRATORE Herpes simplex 1+2 IgM	010	2025-02	1
C2221	CONTROL SERUM Herpes simplex 1+2 IgM	010	2025-02	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	655	≥ 100
Control Serum (INDEX)	2,1	1,5-2,8
Negative Control (INDEX)	0,0	$<0,7$


CALIBRATOR

1 5 6 0 1 0 6 0 0 0 0 0 4 7 0 0 0 0

Label checked with barcode reader

YES ☒

Operator Signature *[Signature]*




CONTROL SERUM

1 5 6 0 1 0 6 1 5 2 8 0 0 0 0 0 0 1

Label checked with barcode reader

YES ☒

Operator Signature *[Signature]*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 19.03.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.