

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME HERPES SIMPLEX 1+2 IgG -CHORUS

Code	81020		
Lot No.	004	2025-04	2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D2212	Herpes simplex 1+2 IgG	004	2025-04	6X6
B2213	CALIBRATORE Herpes simplex 1+2 IgG	004	2025-04	1
C2212	CONTROL SERUM Herpes simplex 1+2 IgG	004	2025-04	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	300	≥ 100
Control Serum (INDEX)	3,3	2,3-4,3
Negative Control (INDEX)	0,0	$<0,7$

CALIBRATOR

0	0	4	0	0	4	7	0	0	0	0	0	3	1	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

0	0	4	0	0	4	7	2	3	4	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 04/04/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.