

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME HERPES SIMPLEX 1 IgG RECOMBINANT - CHORUS

Code 81023

Lot No. 016

2025-03

2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D8681	Herpes simplex 1 IgG Recombinant	016	2025-03	6x6
B8681	CALIBRATORE Herpes simplex 1 IgG Recombinant	016	2025-03	1
C8681	CONTROL SERUM Herpes simplex 1 IgG Recombinant	016	2025-03	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	454	≥ 100
CONTROL SERUM (INDEX)	2,8	1,6-3,0
NEGATIVE CONTROL (INDEX)	0,0	$< 0,7$

CALIBRATOR

2	5	8	0	1	6	6	0	0	0	0	0	4	3	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM

2	5	8	0	1	6	6	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE

11.06.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.