

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME SM - CHORUS

Code 86016

Lot No. 016

2025-06

2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D3161	SM	016	2025-06	6X6
B3161	CALIBRATORE SM	016	2025-06	1
C3161	CONTROL SERUM SM	016	2025-06	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	540	≥ 100
Control Serum (AU/mL)	45,5	35,0-66,2

CALIBRATOR

1	9	2	0	1	6	7	2	1	2	2	4	5	0	3	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *MB*




CONTROL SERUM

1	9	2	0	1	6	7	1	4	2	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *MB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Aguc*

DATE

10.06.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.