

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME LKM-1 CHORUS

Code 86080

Lot No. 016

2025-03

2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D3641	LKM-1	016	2025-03	6X6
B3641	CALIBRATORE LKM-1	016	2025-03	1
C3641	CONTROL SERUM LKM-1	016	2025-03	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	680	≥ 100
Control Serum (AU/mL)	42,2	22,8-60,7

CALIBRATOR

7	0	3	0	1	6	6	4	1	9	3	3	6	0	5	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature HB



CONTROL SERUM

7	0	3	0	1	6	6	1	2	2	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature HB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 12/04/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.