



QUALITY CONTROL CERTIFICATE

KIT NAME HERPES SIMPLEX 1+2 IgM -CHORUS

Code 81021

Lot No. 016

2025-03

2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D2222	Herpes simplex 1+2 IgM	016	2025-03	6x6
B2222	CALIBRATORE Herpes simplex 1+2 IgM	016	2025-03	1
C2222	CONTROL SERUM Herpes simplex 1+2 IgM	016	2025-03	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	549	≥ 100
Control Serum (INDEX)	2,2	1,4-2,6
Negative Control (INDEX)	0,0	$<0,7$

CALIBRATOR

1 5 6 0 1 6 6 0 0 0 0 0 5 0 0 0 0 0

Label checked with barcode reader YES ☒ Operator Signature *FB*



CONTROL SERUM

1 5 6 0 1 6 6 1 4 2 6 0 0 0 0 0 0 0 1

Label checked with barcode reader YES ☒ Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 10/04/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.