

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME **HELICOBACTER PYLORI Ag - CHORUS**

Code	81063		
Lot No.	025	2025-04	2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D6960	Helicobacter pylori Ag	025	2025-04	2X6
B6960	CALIBRATORE Helicobacter pylori Ag	025	2025-06	2
C6960	CONTROL SERUM Helicobacter pylori Ag	025	2025-06	3
G6960	H. pylori Ag Stool Diluent/Negative Control	D1061700	2025-05	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	312	≥ 100
Control + (Index)	3,7	2,7-5,0
Stool Diluent/Negative Control (Index)	0,3	<0,5

CALIBRATOR

9	3	2	0	2	5	6	0	0	0	0	0	2	6	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

9	3	2	0	2	5	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



**This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 15/04/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.