

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME tTg-G - CHORUS

Code	86060		
Lot No.	016	2025-03	2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D3451	tTg-G	016	2025-03	6X6
B3451	CALIBRATORE tTg-G	016	2025-03	1
C3451	CONTROL SERUM tTg-G	016	2025-03	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	518	≥ 100
Control Serum (AU/mL)	48,6	39,2-78,2

CALIBRATOR

3	4	0	0	1	6	6	1	1	3	2	6	4	2	3	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature *MB*



CONTROL SERUM

3	4	0	0	1	6	6	1	7	3	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *MB*



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *MB*

DATE 18.06.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.