

DIESSE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SI) – Italia – Via delle Rose 10
 Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
 20144 Milano (Italy) Via A. Solari 19 sc. 6
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

CERTIFICATO DI CONTROLLO QUALITÀ/QUALITY CONTROL CERTIFICATE

PRODOTTO/PRODUCT ESR CONTROL CUBE

REF: 10435 (4x9 ml)
 10436 (2x9 ml)

REAGENTI/REAGENTS

	Kit	Level 1	Level 2
Lot:	503	503	503
	2019-01	2019-01	2019-01
	2020-07	2020-07	2020-07

RISULTATI/RESULTS

	Level 1		Level 2		Temp.
	Risultato/ Result	Range	Risultato/ Result	Range	
VES-Matic Cube 30	7	1-15	67	42-92	OFF
VES-Matic Cube 80/VES-Matic Cube 200	6	1-14	48	23-73	OFF
MINI-CUBE/Cube 30 touch	8	1-16	53	28-78	ON

Data/Date 15/02/2019

Testato da/Tested by: Katia Tanzini

Approvato da/Approved by: Alessandra Cianchi (Responsabile CQ/QC Head)

Katia Tanzini
Alessandra Cianchi