

DIESE DIAGNOSTICA SENESE SpA53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690

Sede Legale e Amministrazione

20144 Milano (Italy) Via A. Solari 19 sc. 6

Tel. +39-02-4859121 r.a. Fax +39-02-48008530

CERTIFICATO DI CONTROLLO QUALITÀ/QUALITY CONTROL CERTIFICATE**PRODOTTO/PRODUCT**

ESR CONTROL CUBE

REF:

10435 (4x9 ml)

10436 (2x9 ml)

REAGENTI/REAGENTS

	Kit	Level 1	Level 2
Lot:	515	515	515
	2019-04	2019-04	2019-04
	2020-10	2020-10	2020-10

RISULTATI/RESULTS

	Level 1		Level 2		Temp.
	Risultato/ Result	Range	Risultato/ Result	Range	
VES-Matic Cube 30	7	1-15	69	44-94	OFF
VES-Matic Cube 80/VES-Matic Cube 200	6	1-14	49	24-74	OFF
MINI-CUBE/Cube 30 touch	8	1-16	57	32-82	ON

Data/Date31/05/2019**Testato da/Tested by:**

Katia Tanzini
Approvato da/Approved by:

Alessandra Cianchi (Responsabile CQ/QC Head)

DIESSSE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SI) – Italia – Via delle Rose 10
 Tel +39-0577-587111 Fax +39-0577-318690





Sede Legale e Amministrazione
 20144 Milano (Italy) Via A. Solari 19 sc. 6
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

CERTIFICATO DI CONTROLLO QUALITÀ/QUALITY CONTROL CERTIFICATE

PRODOTTO/PRODUCT ESR CONTROL

REF: 10430 (4x9 ml)
 10434 (2x9 ml)

REAGENTI/REAGENTS

	Kit	Level 1	Level 2
Lot:	050	050	050
	2019-04	2019-04	2019-04
	2020-10	2020-10	2020-10

RISULTATI/RESULTS

	Level 1		Level 2		Temp.
	Risultato/ Result	Range	Risultato/ Result	Range	
MINI-VES	11	3-19	90	65-115	OFF
VES-Matic Easy	6	1-14	75	50-100	ON
VES-Matic 10	4	1-14	73	48-98	ON
VES-Matic 20	7	1-15	78	53-103	ON
VES-Matic 30	6	1-14	81	56-106	ON
VES-Matic 60	8	1-16	88	63-113	ON
VES-Matic 30 "Fast method"	35	17-53	106	81-131	ON

Data/Date

31/05/2019

Testato da/Tested by:


 Katia Tanzini

Approvato da/Approved by:


 Alessandra Cianchi (Responsabile CQ/QC Head)