

DIESSE DIAGNOSTICA SENESE SpA53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690

Sede Legale e Amministrazione

20144 Milano (Italy) Via A. Solari 19 sc. 6

Tel. +39-02-4859121 r.a. Fax +39-02-48008530

CERTIFICATO DI CONTROLLO QUALITÀ/QUALITY CONTROL CERTIFICATE**PRODOTTO/PRODUCT**



ESR CONTROL CUBE

REF:

10435 (4x9 ml)

10436 (2x9 ml)

REAGENTI/REAGENTS

	Kit	Level 1	Level 2
Lot:	515	515	515
	2019-04	2019-04	2019-04
	2020-10	2020-10	2020-10

RISULTATI/RESULTS

	Level 1		Level 2		Temp.
	Risultato/ Result	Range	Risultato/ Result	Range	
VES-Matic Cube 30	7	1-15	69	44-94	OFF
VES-Matic Cube 80/VES-Matic Cube 200	6	1-14	49	24-74	OFF
MINI-CUBE/Cube 30 touch	8	1-16	57	32-82	ON

Data/Date

31/05/2019

Testato da/Tested by:

 Katia Tanzini
Approvato da/Approved by:

 Alessandra Cianchi (Responsabile CQ/QC Head)