

**DIESE DIAGNOSTICA SENESE SpA**  
 Via delle Rose 10, 53035 Monteriggioni (SI) - Italy  
 Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione  
 Via Solari 19, sc.6 20144 Milano - Italy  
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

**QUALITY CONTROL CERTIFICATE**

**KIT NAME** DIESE STREPTO SLIDE ESTRAENTI

**Code** 0217

**LOT.NO 107** **EXPIRY** 2021-03

**REAGENTS**

CODE	LABEL	DESCRIPTION	LOT No.	EXP	NO
PF0217A	B0331	REAGENT 1	041	2022-01	1
PF0217B	C0331	REAGENT 2	037	2022-01	1
PF0217C	D0330	REAGENT 3	042	2022-01	2
PF01217	E0330	LYTIC ENZYME	034	2021-03	2
PF01210	F0330	POSITIVE CONTROL	064	2021-06	1

**CONTROLS**

	FOUND RESULTS	EXPECTED RESULTS
Test with Streptococcus A group	POSITIVE	POSITIVE
Test with Streptococcus F group	POSITIVE	POSITIVE

The test has been carried out in accordance with the package insert

**TESTED BY:** Silvia Sardelli

**APPROVED BY:** (RESPONSIBLE HEAD, Quality Control Reagents)

**DATE** 09-03-2020

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**  
 This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.