

DIESE DIAGNOSTICA SENESE SpA
 Via delle Rose 10, 53035 Monteriggioni (SI), Italy
 Tel. ++39-0577-587111 fax ++39-0577-318690

Sede Legale e Amministrazione
 Via A. Solari 19 20144 Milano Italy
 Tel. ++39-02-4859121 r.a. fax ++39-02-48008530

CERTIFICATE OF ANALYSIS

CODE	PRODUCT NAME	LOT	EXPIRY DATE	PREP DATE
10439	SediCheck Cube 30 4 x 3 ml	022	2021-01	2020-01

Level I (NORMAL Level) lot 605	Test conditions	VES 1st hour	Range of acceptability
VES-MATIC CUBE 30	Temperature OFF	5	1-14
MINI-CUBE	Temperature ON		
Cube 30 Touch	Temperature ON		

Level II (NORMAL Level) lot 605	Test conditions	VES 1st hour	Range of acceptability
VES-MATIC CUBE 30	Temperature OFF	55	30-80
MINI-CUBE	Temperature ON		
Cube 30 Touch	Temperature ON		

Date: 10-03-2020



Signature (Head Quality Control)