





QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS INFLUENZA A CONTROL SERUM

Code	85872				
Lot. No	090		2021-06		2020-06

REAGENTS

CODE	LABEL	DESCRIPTION	Lot No.	EXP	NO
PF 90872	B2950	CHORUS INFLUENZA A	090	2021-06	1

CONTROLS

TESTED WITH:	FOUND RESULTS	EXPECTED RESULTS
CHORUS INFLUENZA A LOT 616	1/32	1/32 ± 1 Dil.

The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo 

APPROVED BY:  (RESPONSIBLE HEAD, Quality Control Reagents)

DATE 24/06/2020

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.