

DIESSE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SI) – Italia – Via delle Rose 10
 Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
 20144 Milano (Italy) Via A. Solari 19 sc. 6
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

CERTIFICATO DI CONTROLLO QUALITÀ/QUALITY CONTROL CERTIFICATE

PRODOTTO/PRODUCT ESR CONTROL CUBE

REF: 10435 (4x9 ml)
 10436 (2x9 ml)

REAGENTI/REAGENTS

	Kit	Level 1	Level 2
Lot:	621	621	621
	2020-05	2020-05	2020-05
	2021-11	2021-11	2021-11

RISULTATI/RESULTS

	Level 1		Level 2		Temp.
	Risultato/ Result	Range	Risultato/ Result	Range	
VES-Matic Cube 30	7	1-15	83	58-108	OFF
VES-Matic Cube 80/VES-Matic Cube 200	7	1-15	62	37-87	OFF
MINI-CUBE/Cube 30 touch	9	1-17	72	47-97	ON

Data/Date

05-08-2020

Testato da/Tested by:

Katia Tanzini

Approvato da/Approved by:

Alessandra Cianchi (Responsabile CQ/QC Head)

Katia Tanzini
Alessandra Cianchi