

**DIESE DIAGNOSTICA SENESE SpA**  
53035 Monteriggioni (SI) – Italia – Via delle Rose 10  
Tel +39-0577-587111 Fax +39-0577-318690





Sede Legale e Amministrazione  
20144 Milano (Italy) Via A. Solari 19 sc. 6  
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

**CERTIFICATO DI CONTROLLO QUALITÀ/QUALITY CONTROL CERTIFICATE**

**PRODOTTO/PRODUCT** ESR CONTROL CUBE

**REF:** 10435 (4x9 ml)  
10436 (2x9 ml)

**REAGENTI/REAGENTS**

	Kit	Level 1	Level 2
Lot:	704	704	704
	2021-01	2021-01	2021-01
	2022-07	2022-07	2022-07

**RISULTATI/RESULTS**

	Level 1		Level 2		Temp.
	Risultato/ Result	Range	Risultato/ Result	Range	
VES-Matic Cube 30	7	1-15	73	48-98	OFF
VES-Matic Cube 80/VES-Matic Cube 200/Ves-Matic Cube Track	5	1-14	54	29-79	OFF
MINI-CUBE/Cube 30 touch	6	1-14	59	34-84	ON

**Data/Date**

22-02-2021

**Testato da/Tested by:**

Katia Tanzini

*Katanzini*  
*Alessandra*

**Approvato da/Approved by:**

Alessandra Cianchi

(Responsabile CQ/QC Head)