



EU DECLARATION OF CONFORMITY

MANUFACTURER:	DIESSE DIAGNOSTICA SENESE SPA STRADA DEI LAGHI 39 53035 MONTERIGGIONI (SI), ITALY
SINGLE REGISTRATION NUMBER	IT-MF-000013311
AUTHORIZED REPRESENTATIVE:	NOT APPLICABLE
PRODUCT: CODE:	VES-MATIC 5 10360
INTENDED PURPOSE:	<p>The VES-MATIC 5 is an automated instrument for the quantitative Erythrocyte Sedimentation Rate (ESR) determination, measured using modified Westergren method on venous whole blood anticoagulated with EDTA. ESR is a non-specific parameter of an inflammatory status, used as an aid for the monitoring of the physiological or pathological state of the patient.</p> <p>The instrument is to be used only by professional laboratory users.</p>
BASIC UDI-DI	803389132VESMATIC5008M
UDI-DI	08033891324230
RISK CLASS:	CLASS B
CLASSIFICATION RULE:	RULE 6
CONFORMITY ASSESSMENT ROUTE:	ANNEX IX

WE HEREWITH DECLARE THAT THE ABOVE MENTIONED PRODUCT IS IN CONFORMITY WITH **REGULATION (EU) 2017/746** OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL ON IN VITRO DIAGNOSTIC MEDICAL DEVICE AND WITH THE FOLLOWING UNION LEGISLATION: **LOW VOLTAGE DIRECTIVE (2014/35/EU)**, **ELECTROMAGNETIC COMPATIBILITY DIRECTIVE (2014/30/EU)** AND **RESTRICTION OF HAZARDOUS SUBSTANCES IN ELECTRICAL AND ELECTRONIC EQUIPMENT DIRECTIVE – ROHS2 (2011/65/EU)**.

ALL SUPPORTING DOCUMENTATION IS RETAINED UNDER THE PREMISES OF THE MANUFACTURER.

THIS CERTIFICATE WILL LOSE ITS VALIDITY IN THE EVENT OF:

- MODIFICATIONS MADE TO THE MACHINE IN QUESTION WITHOUT OUR AUTHORIZATION
- INCORRECT USE OF THE INSTRUMENT
- TECHNICAL INTERVENTIONS PERFORMED BY UNAUTHORIZED PERSONNEL
- INSTALLATION OF NON-ORIGINAL SPARE PARTS.

REFERENCE TO ANY CS APPLIED:	NOT APPLICABLE
------------------------------	----------------

NOTIFIED BODY:

TÜV SÜD PRODUCT SERVICE GMBH
RIDLERSTAßE 65 – 80339 MUNICH
GERMANY
No. 0123

(EU) CERTIFICATE:

V12 056726 0006 Rev. 01

REVISION:

0

PLACE, DATE OF ISSUE:

MONTERIGGIONI, 20 DECEMBER 2023

EXPIRY DATE:

2027-06-26

THE PRESENT EU DECLARATION OF CONFORMITY IS ISSUED UNDER THE SOLE AND EXCLUSIVE RESPONSIBILITY OF THE MANUFACTURER.

SIGNATURE:



MARIA CLAUDIA ALCARO
PERSON RESPONSIBLE FOR THE REGULATORY
COMPLIANCE

THIS IS THE CERTIFIED COPY OF THE ORIGINAL DOCUMENT STORED IN ARCHIVE OF DIESSE DIAGNOSTICA SENESE SPA

ISSUED: MONTERIGGIONI, 2023-12-20



MAGDALENA STOCZKO
REGULATORY SUPERVISOR