

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SIENA) – Italy – Strada dei Laghi, 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME: ANTI-CCP - CHORUS

Code	86094		
Lot No.	743	2022-06	2021-07

REAGENTS

Label	Description	Lot.	Exp	No
D8670	ANTI-CCP	743	2022-06	6X6
B8670	CALIBRATORE ANTI-CCP	743	2022-07	1
C8670	CONTROL SERUM ANTI-CCP	743	2022-07	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	360	≥ 100
Control Serum (AU/mL)	56,7	38,3-78,9

CALIBRATOR


2	5	6	7	4	3	6	4	2	2	4	1	4	0	6	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM


2	5	6	7	4	3	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents

DATE

23/07/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.