

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME ZIKA IgM CAPTURE -CHORUS

Code	81161		
Lot No.	826	2023-04	2022-05

REAGENTS

Label	Description	Lot.	Exp	No
D4290	Zika IgM Capture	826	2023-04	6X6
B4290	Calibratore Zika IgM Capture	826	2023-04	1
C4290	Control Zika IgM Capture	826	2023-04	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	354	≥ 100
Control (Index)	3,3	2,2-4,1


CALIBRATOR

9	8	7	8	2	6	6	0	0	0	0	0	3	2	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



CONTROL SERUM

9	8	7	8	2	6	6	2	2	4	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



This batch has been manufactured and checked according to quality requirements.
Hence, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 18.05.2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.