

DIESE DIAGNOSTICA GENESE SPA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: DEAMIDATED GLIADIN-G - CHORUS

Code	85100		
Lot No.	827	2023-04	2022-04

REAGENTS

Label	Description	Lot.	Exp	No
D8660	Deamidated Gliadin-G	827	2023-04	6X6
B8660	CALIBRATORE Deamidated Gliadin-G	827	2023-04	1
C8660	CONTROL SERUM Deamidated Gliadin-G	827	2023-04	1

CONTROLS

Found results	Expected results
---------------	------------------


Calibrator (mOD)	302	≥ 100
Control Serum (AU/mL)	98,3	64,3-196,0

CALIBRATOR

2	5	5	8	2	7	6	2	2	1	3	5	2	8	6	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES


Operator Signature *Sf*

CONTROL SERUM

2	5	5	8	2	7	6	2	5	4	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature *Sf*

The test has been carried out in accordance with the package insert

TESTED BY: Simone Balocchi

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents.

DATE

18.05.22

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.