

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: TG Extended Range

Code	86110		
Lot No.	834	2023-05	2022-05

REAGENTS

Label	Description	Lot.	Exp	No
D4050	TG Extended Range	834	2023-05	6X6
B4050	CALIBRATOR TG Extended Range	834	2023-05	1
C4050	CONTROL SERUM TG Extended Range	834	2023-05	1

CONTROLS

Found results	Expected results
---------------	------------------

CALIBRATOR (mOD)	410	≥100
CONTROL SERUM (ng/ml)	3,8	2,0-5,5

CALIBRATOR


4	1	8	8	3	4	6	6	1	3	2	4	3	0	4	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM


4	1	8	8	3	4	6	2	4	4	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 22-05-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.