

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SIENA) – Italy – Strada dei Laghi, 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME: SS-B - CHORUS

Code	86024		
Lot No.	837	2023-05	2022-05

REAGENTS

Label	Description	Lot.	Exp	No
D3201	SS-B	837	2023-05	6X6
B3201	CALIBRATORE SS-B	837	2023-05	1
C3201	CONTROL SERUM SS-B	837	2023-05	1

CONTROLS

Found results	Exp
---------------	-----

Calibrator (mOD)	420	≥ 100
Control Serum (AU/mL)	47,9	33,3-58,3

CALIBRATOR


1	9	4	8	3	7	6	3	1	6	3	4	3	9	5	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM

1	9	4	8	3	7	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 09.08.2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.