

DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: 88000-T9 OLEA EUROPEA

Code	88000-T9		
Lot No.	834	2023-05	2022-06

REAGENTS

Label	Description	Lot	Exp	No
D9921	88000-T9 OLEA EUROPEA	834	2023-05	2X6
B9920	CALIBRATORE 88000-T9 OLEA EUROPEA	834	2023-06	1
C9920	CONTROL SERUM 88000-T9 OLEA EUROPEA	834	2023-06	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	322	≥ 100
Control Serum (KU/L)	13,5	11,0 - 35,1

CALIBRATOR


6	0	9	8	3	4	6	5	2	0	5	3	2	2	8	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *YSH*



CONTROL SERUM

6	0	9	8	3	4	6	3	7	6	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *YSH*



The test has been carried out in accordance with the package insert

TESTED BY: Irene Maccagli

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents

DATE 28/06/2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.