

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: tTg-A - CHORUS

Code	88058		
Lot No.	844	2023-06	2022-06

REAGENTS

Label	Description	Lot.	Exp	No
D3431	tTg-A	844	2023-06	6X6
B3431	CALIBRATORE tTg-A	844	2023-06	1
C3431	CONTROL SERUM tTg-A	844	2023-06	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	320	≥ 100
Control Serum (AU/mL)	40,3	30,0-63,3

CALIBRATOR


3	3	9	8	4	4	6	2	2	0	3	5	3	7	4	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *Chen*



CONTROL SERUM


3	3	9	8	4	4	6	2	0	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *Chen*



The test has been carried out in accordance with the package insert

TESTED BY: Simone Balocchi

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 12.07.2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.