

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SIENA) – Italy – Strada del Laghi, 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME: CARDIOLIPIN-G - CHORUS

Code	86046		
Lot No.	845	 2023-06	 2022-06

REAGENTS

Label	Description	Lot.	Exp	No
D3311	CARDIOLIPIN-G	845	2023-06	6X6
B3311	CALIBRATORE CARDIOLIPIN-G	845	2023-06	1
C3311	CONTROL SERUM CARDIOLIPIN-G	845	2023-06	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	360	≥ 100
Control Serum (GPLU/mL)	29,1	21,4-40,6

CALIBRATOR


3	3	5	8	4	5	6	3	2	4	4	1	4	3	6	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *[Signature]*



CONTROL SERUM

3	3	5	8	4	5	6	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *[Signature]*



The test has been carried out in accordance with the package Insert

TESTED BY: Simone Balocchi

[Signature]

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

[Signature]

DATE

26.07.2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.