

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SIENA) – Italy – Strada dei Laghi, 39
Tel. +39 0577319554 FAX +39 0577319020



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME: SM - CHORUS

Code	86016
Lot No.	819
	2023-06
	2022-03

REAGENTS

Label	Description	Lot.	Exp	No
D3161	SM	819	2023-06	6X6
B3161	CALIBRATORE SM	819	2023-06	1
C3161	CONTROL SERUM SM	819	2023-06	1

CONTROLS


	Found results	Expected results
Calibrator (mOD)	291	≥ 100
Control Serum (AU/mL)	72,1	50,0-79,3

CALIBRATOR

1	9	2	8	1	9	7	1	1	3	2	2	3	6	5	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature *SB*

CONTROL SERUM

1	9	2	8	1	9	7	1	9	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature *SB*

The test has been carried out in accordance with the package insert

TESTED BY: Simone Baiocchi

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 29.07.2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other Infectious agents are absent, reagents should be handled with maximum attention.