

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME ANTI-CCP - CHORUS

Code 86094

Lot No. 839



2023-05



2022-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D8671	ANTI-CCP	839	2023-05	6x6
B8671	CALIBRATORE ANTI-CCP	839	2023-05	1
C8671	CONTROL SERUM ANTI-CCP	839	2023-05	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	508	≥ 100
Control Serum (AU/mL)	71,0	54,5-140,0

CALIBRATOR

2	5	6	8	3	9	6	5	1	6	2	7	4	2	4	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *Se*




CONTROL SERUM

2	5	6	8	3	9	6	1	7	3	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *Se*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Aliboni*

DATE 10/08/2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.