

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: CALPROTECTIN - CHORUS

Code 86600		
Lot No. 855	2023-07	2022-07

REAGENTS

Label	Description	Lot.	Exp	No
D9500	CALPROTECTIN	855	2023-07	2X6
B9500	CALIBRATORE CALPROTECTIN	855	2023-11	1
C9500	CONTROL CALPROTECTIN	855	2023-11	1
G9502	SAMPLE DILUENT	855	2024-07	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	420	≥ 100
Control (mg/kg)	139.0	71.4-194.4

CALIBRATOR


9	0	0	8	5	5	6	2	1	7	2	1	5	3	3	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *Simone Baiocchi*



CONTROL

9	0	0	8	5	5	6	1	3	2	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *Simone Baiocchi*



The test has been carried out in accordance with the package insert

TESTED BY: Simone Baiocchi

Simone Baiocchi

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents

Simone Baiocchi

DATE

25/08/2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.