

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI) – Italia – Via delle Rose 10  
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione  
20144 Milano (Italy) Via A. Solari 19 sc. 6  
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

Pag 1/1

**PRODUCT NAME:** VACU-TEC

**Code** 10208

**Lot No.** 2215



2023-06



2022-08

**Reference** IO-09/088

**ON LINE QUALITY CONTROLS** (model CDP not enclosed)

**QC Passed**

- ☒ sodium citrate solution concentration
- ☒ sodium citrate volume
- ☒ vacuum level (\*)
- ☒ correct placement of the cap
- ☒ no extra body inside the product

✓

✓

✓

✓

✓

(\*) Only for VACU-CODE and VACU-TEC tubes

✓

**PACKAGING QUALITY CONTROL** (model CDP not enclosed)

**QC Passed**

- ☒ control of the number of the product for each pack

✓

**QUALITY CONTROLS ON FINISHED PRODUCT** (model CDP not enclosed)

**QC Passed**

- ☒ pack integrity
- ☒ data printed on the product label

✓

✓

**QC Passed**

N/A

**Irradiation treatment Certificate n°** .....

✓

✓

**QC Passed**

N/A

**Gamma rays treatment process in compliance with the specifications of the supply Agreement**

✓

✓

Document number:

ODV N.A.

*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

**APPROVED**

RESPONSIBLE HEAD, Quality Control Reagents.....

**DATE** ...05/08/2022