

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SIENA) – Italy – Strada dei Laghi, 39
Tel. +39 0577319554 FAX +39 0577319020



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME: CARDIOLIPIN-G - CHORUS

Code	86046		
Lot No.	847	2023-06	2022-06

REAGENTS

bel	Description	Lot.	Exp	No
D3311	CARDIOLIPIN-G	847	2023-06	6X6
B3311	CALIBRATORE CARDIOLIPIN-G	847	2023-06	1
C3311	CONTROL SERUM CARDIOLIPIN-G	847	2023-06	1

CONTROLS


	Found results	Expected results
Calibrator (mOD)	511	≥ 100
Control Serum (GPLU/mL)	33,1	21,4-40,6

CALIBRATOR

3	3	5	8	4	7	6	3	2	4	4	1	4	3	6	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES


Operator Signature *SB*

CONTROL SERUM

3	3	5	8	4	7	6	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature *SB*

The test has been carried out in accordance with the package insert

TESTED BY: Simone Balocchi

Simone Balocchi

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents

DATE

29.08.2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.