

DIESSE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SIENA) – Italy – Strada dei Laghi, 39
 Tel. +39 0577319554 FAX +39 0577319020



Sede Legale e Amministrazione
 Via A. Solari, 19 20144 Milano - Italy

QUALITY CONTROL CERTIFICATE

KIT NAME ENZYWELL TOXOPLASMA IgM

Code	91041		
LOT. NO.	314-A	2023-11	2022-08

REAGENTS

CODE	LABEL	DESCRIPTION	LOT No.	EXP	NO
PF93000	006330	MICROPLATE	663	2023-12	1
PF91540	006320	CONJUGATE	220	2023-11	1
PF93900/TXM	C5100	NEGATIVE CONTROL	761	2024-02	1
PF92042	005370	POSITIVE CONTROL	201	2024-02	1
PF91841	005120	CUT-OFF CONTROL	199	2024-02	1
PF91641	005130	ANTIGEN (3,0 mL)	204	2025-02	6
PF93611	009660	DILUENT 2	225	2023-12	1
PF93603	1035	WASHING BUFFER 10x	451-1	2024-01	1
PF93602	1032/RISK0	STOP SOLUTION	215	2024-08	1
PF93619	1034	SUBSTRATE-HS	365	2023-11	1

CONTROLS

	FOUND RESULTS	EXPECTED RESULTS
NEGATIVE CONTROL	0.175	< 0,6 CUT-OFF
CUT-OFF CONTROL	0.380	≥ 0.200
POSITIVE CONTROL	2.458	> 1,5 CUT-OFF
CV %	9%	≤ 15%

The test has been carried out in accordance with the package insert

TESTED BY: Katia Tanzini

APPROVED BY: RESPONSIBLE HEAD, Quality Control Reagents

DATE 16.09.2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.



CERTIFICATE OF ANALYSIS

KIT: TOXOPLASMA IgM Code 91041
 LOT: 314 Code 91179
 DATE OF TEST: 09/09/2022

METHOD: ACCORDING TO THE PACKAGE INSERT

1. SOLID PHASE		FOUND	EXPECTED
A) UNIFORMITY BETWEEN PLATES			
1st PLATE	OD 450 nm	1.678	
	CV%	12%	< 20%
2nd PLATE	OD 450 nm	1.859	
	CV%	8%	< 20%
B) UNIFORMITY WITHIN PLATE			
	OD 450 nm	1.804	
	CV%	8%	< 15%
2. BULK CONTROL date: 01/09/2022			
A) Average OD of Cut Off Control	OD 450 nm	0.425	≥ 0.200
B) Index of Negative and Positive Sera		accepted	According to specification
C) Sensitivity with working calibrators		accepted	According to specification
3. FINAL CONTROL date: 09/09/2022			
A) Blank	OD 450 nm	0.084	≤ 0.100
B) Average OD of Negative Control	OD 450 nm	0.175	< 0,6Cut Off
C) Average OD of Cut Off Control	OD 450 nm	0.380	≥ 0.200
D) Average OD of Positive Control	OD 450 nm	2.458	> 1,5Cut Off
E) HOMOGENEITY OF CUT OFF	No.	6	
	OD 450 nm	0.464	
	CV%	9%	≤ 15%
F) PRECISION OF FILLING PROCESS			

☒ SATISFACTORY ☐ UNSATISFACTORY

FINAL RESULT: **ACCEPTED**

Date: 12.09.2022

Signed: 
 Released: Q.C.Manager

Aggiornamento 07.2012