

DIESSE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SIENA) – Italy – Strada dei Laghi, 39
 Tel. +39 0577319554 FAX +39 0577319020



Sede Legale e Amministrazione
 Via A. Solari, 19 20144 Milano - Italy

QUALITY CONTROL CERTIFICATE

KIT NAME ENZYWELL TREPONEMA IgG

Code 91050

LOT.NO 140 EXPIRY 2023-11

REAGENTS

CODE	LABEL	DESCRIPTION	LOT No.	EXP	NO
PF93050	006820	MICROPLATE	412	2024-01	1
PF93505	006810	CONJUGATE	108	2023-11	1
PF93910	1036	NEGATIVE CONTROL	174	2024-02	1
PF92050	005040	POSITIVE CONTROL	091	2024-02	1
PF91850	005060	CUT-OFF CONTROL	101	2024-02	1
PF93619	1034	SUBSTRATE-HS	365	2023-11	1
PF93621	006030	DILUENT	161	2024-01	1
PF93603	1035	WASHING BUFFER 10x	451-2	2024-01	1
PF93602	1032/RISK0	STOP SOLUTION	215	2024-08	1

CONTROLS

IU/mL	FOUND RESULTS	EXPECTED RESULTS
NEGATIVE CONTROL 0	0.018	$\leq 0,6$ CUT-OFF
CUT-OFF CONTROL 0,25	0.413	≥ 0.200
POSITIVE CONTROL 2	2.462	$\geq 1,5$ CUT-OFF
CV %	11%	$\leq 15\%$

The test has been carried out in accordance with the package insert
 The IU/mL are calculated against the 1st International Standard Preparation 1958.

TESTED BY: Silvia Sardelli

APPROVED BY: (RESPONSIBLE HEAD, Quality Control Reagents)

DATE: 22.06.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.



CERTIFICATE OF ANALYSIS

KIT: **TREPONEMA IgG** Code 91050
 LOT: **140**
 DATE OF TEST: **20/09/2022**

METHOD: ACCORDING TO THE PACKAGE INSERT

1. MICROPLATE CONTROL	FOUND	EXPECTED
A) UNIFORMITY BETWEEN PLATES		
1st PLATE	OD 450 nm 2.365	
	CV% 3%	≤20%
B) UNIFORMITY WITHIN PLATE		
	OD 450nm 2.258	
	CV% 6%	≤15%
2. BULK CONTROL date: 13/09/2022		
A) Average OD of Cut Off Control	OD 450nm 0.460	≥ 0.200
B) Average OD of 5 Negative Sera	accepted	According to specification
C) Average OD of 3 Positive Sera	accepted	According to specification
D) Sensitivity with working calibrators	accepted	According to specification
3. FINAL CONTROL date: 20/09/2022		
A) Blank	OD 450 nm 0.038	≤ 0.100
B) Average OD of Negative Control	OD 450 nm 0.018	< 0.6xCut Off
C) Average OD of Cut Off Control	OD 450 nm 0.413	≥ 0.200
D) Average OD of Positive Control	OD 450 nm 2.462	> 1.5 x Cut Off
E) HOMOGENEITY OF CUT-OFF	N° Replicate 6	
	OD 450 nm 0.451	
	CV% 11%	≤ 15%
F) PRECISION OF FILLING PROCESS		

SATISFACTORY

☒

UNSATISFACTORY

FINAL RESULT: **ACCEPTED**

Date: *20.09.2022*

Signed: *[Signature]*
 Released: Q.C.Manager

Aggiornamento 08.2012