

QUALITY CONTROL CERTIFICATE

KIT NAME GLIADIN-A - CHORUS

Code	86054
Lot No.	861
	2023-08
	2022-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D3391	Gliadin-A	861	2023-08	6X6
B3392	CALIBRATORE Gliadin-A	861	2023-09	1
C3392	CONTROL SERUM Gliadin-A	861	2023-09	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

CALIBRATOR (mOD)	370	> 100
CONTROL SERUM (AU/ml)	38,9	30,2-60,7

CALIBRATOR

5	0	1	8	6	1	6	4	4	0	5	4	3	0	6	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature *89*



CONTROL SERUM

5	0	1	8	6	1	6	2	3	4	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *88*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE *07-10-2022*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.