

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME ANTI-CCP - CHORUS

Code	86094		
Lot No.	870	2023-09	2022-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D8671	ANTI-CCP	870	2023-09	6x6
B8671	CALIBRATORE ANTI-CCP	870	2023-09	1
C8671	CONTROL SERUM ANTI-CCP	870	2023-09	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	523	≥ 100
Control Serum (AU/mL)	58,9	44,2-93,8

CALIBRATOR

2	5	6	8	7	0	6	5	2	2	3	8	3	6	5	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

2	5	6	8	7	0	6	2	0	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



**This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 20.10.2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.