

DIESE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

## QUALITY CONTROL CERTIFICATE

KIT NAME **tTg-G - CHORUS**

|         |       |         |         |
|---------|-------|---------|---------|
| Code    | 86060 |         |         |
| Lot No. | 878   | 2023-10 | 2022-10 |

### REAGENTS

| Label | Description         | Lot. | Exp.    | NO  |
|-------|---------------------|------|---------|-----|
| D3451 | tTg-G               | 878  | 2023-10 | 6X6 |
| B3451 | CALIBRATORE tTg-G   | 878  | 2023-10 | 1   |
| C3451 | CONTROL SERUM tTg-G | 878  | 2023-10 | 1   |

### CONTROLS


|  | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

|                       |      |            |
|-----------------------|------|------------|
| Calibrator (mOD)      | 380  | $\geq 100$ |
| Control Serum (AU/mL) | 37,4 | 26,3-54,0  |

### CALIBRATOR

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 3 | 4 | 0 | 8 | 7 | 8 | 6 | 1 | 1 | 8 | 3 | 3 | 4 | 4 | 4 | 0 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|


Label checked with barcode reader YES ☒ Operator Signature



### CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 3 | 4 | 0 | 8 | 7 | 8 | 6 | 1 | 6 | 3 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.

DATE 27-10-2022

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.