

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME SS-A - CHORUS

Code 86018

Lot No. 855



2023-10



2022-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D3181	SS-A	855	2023-10	6X6
B3181	CALIBRATORE SS-A	855	2023-10	1
C3181	CONTROL SERUM SS-A	855	2023-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	320	≥ 100
Control Serum (AU/mL)	49,2	32,0-65,0

CALIBRATOR

1	9	3	8	5	5	7	2	1	7	3	2	3	9	4	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

88



CONTROL SERUM

1	9	3	8	5	5	7	1	8	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

88



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

88

DATE

04.11.2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.