

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CALPROTECTIN - CHORUS

Code	86600		
Lot No.	861	2023-08	2022-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D9500	CALPROTECTIN	861	2023-08	2 X 6
B9500	CALIBRATORE CALPROTECTIN	861	2023-11	1
C9500	CONTROL CALPROTECTIN	861	2023-11	1
G9502	SAMPLE DILUENT	861	2024-10	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	460	≥ 100
Control (mg/kg)	128,8	80,0-212,5

CALIBRATOR

9	0	0	8	6	1	6	2	2	0	2	4	4	5	3	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

9	0	0	8	6	1	6	1	6	2	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 02.11.22

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.