

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SIENA) – Italy – Strada del Laghi, 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME: AMA-M2 - CHORUS

Code	86084		
Lot No.	861	2023-11	2022-08

REAGENTS

Label	Description	Lot.	Exp	No
D3651	AMA-M2	861	2023-11	6X6
B3651	CALIBRATORE AMA-M2	861	2023-11	1
C3651	CONTROL SERUM AMA-M2	861	2023-11	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	429	≥ 100
Control Serum (AU/mL)	46,4	39,2-137,5

CALIBRATOR

7	0	5	8	6	1	7	3	2	3	3	1	3	9	4	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

[Signature]

705861732331394700	
AMA-M2	
CALIBRATOR	0.176 mL TVD
LOT	861
2023-11	2°C
DIESSE	Monteriggioni Italy

CONTROL SERUM

7	0	5	8	6	1	7	1	9	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

[Signature]

705861719340000001	
AMA-M2	
CONTROL+	0.425 mL TVD
LOT	861
2023-11	2°C
DIESSE	Monteriggioni Italy

The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo *[Signature]*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 07-11-2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.