

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME GLIADIN-G - CHORUS

Code 86056

Lot No. 869



2023-09



2022-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D3411	Gliadin-G	869	2023-09	6X6
B3411	CALIBRATORE Gliadin-G	869	2023-10	1
C3411	CONTROL SERUM Gliadin-G	869	2023-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

CALIBRATOR (mOD)	480	≥ 100
CONTROL SERUM (AU/ml)	56,3	37,3-91,7

CALIBRATOR


5	0	2	8	6	9	6	4	2	1	3	3	4	3	5	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature 



CONTROL SERUM


5	0	2	8	6	9	6	1	7	3	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature 



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... 

DATE 

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.