



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME ASCA-A - CHORUS

Code	86062		
Lot No.	878	2023-10	2022-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D3471	ASCA-A	878	2023-10	6X6
B3471	CALIBRATORE ASCA-A	878	2023-10	1
C3471	CONTROL SERUM ASCA-A	878	2023-10	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	284	≥ 100
Control Serum (AU/mL)	112,1	60,7-161,5

CALIBRATOR

5	0	7	8	7	8	6	4	1	9	3	3	3	2	5	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

5	0	7	8	7	8	6	2	2	4	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 16.11.2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.