

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME MPO - CHORUS

|         |       |         |         |
|---------|-------|---------|---------|
| Code    | 86068 |         |         |
| Lot No. | 871   | 2023-09 | 2022-09 |

### REAGENTS

| Label | Description       | Lot. | Exp.    | NO  |
|-------|-------------------|------|---------|-----|
| D3531 | MPO               | 871  | 2023-09 | 6X6 |
| B3531 | CALIBRATORE MPO   | 871  | 2023-09 | 1   |
| C3531 | CONTROL SERUM MPO | 871  | 2023-09 | 1   |

### CONTROLS

|  | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

|                       |      |            |
|-----------------------|------|------------|
| Calibrator (mOD)      | 460  | $\geq 100$ |
| Control Serum (AU/mL) | 66,6 | 48,0-94,7  |

### CALIBRATOR


|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 2 | 5 | 3 | 8 | 7 | 1 | 6 | 1 | 1 | 5 | 2 | 5 | 3 | 4 | 4 | 0 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES



Operator Signature



### CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 2 | 5 | 3 | 8 | 7 | 1 | 6 | 2 | 1 | 3 | 8 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.