

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME DEAMIDATED GLIADIN-A - CHORUS

Code 86098

Lot No. 883



2023-10



2022-11

### REAGENTS

Label	Description	Lot.	Exp.	NO
D8650	Deamidated Gliadin-A	883	2023-10	6X6
B8650	CALIBRATORE Deamidated Gliadin-A	883	2023-10	1
C8650	CONTROL SERUM Deamidated Gliadin-A	883	2023-10	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	473	> 100
Control Serum (AU/mL)	38,5	27,7-81,2

### CALIBRATOR

2	5	4	8	8	3	6	2	2	1	2	9	5	1	4	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

*[Signature]*



### CONTROL SERUM

2	5	4	8	8	3	6	1	4	2	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

*[Signature]*



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

*[Signature]*

DATE 23-11-2022

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.