

DIESE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel, +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME tTg-G - CHORUS

Code 86060

Lot No. 881



2023-10



2022-10

### REAGENTS

Label	Description	Lot.	Exp.	NO
D3451	tTg-G	881	2023-10	6X6
B3451	CALIBRATORE tTg-G	881	2023-10	1
C3451	CONTROL SERUM tTg-G	881	2023-10	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	340	$\geq 100$
Control Serum (AU/mL)	38,9	26,3-54,0

### CALIBRATOR


3	4	0	8	8	1	6	1	1	8	3	3	4	4	4	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature .....



### CONTROL SERUM


3	4	0	8	8	1	6	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature .....



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 24.11.2022

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.