

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS PROMONITOR ANTI-VEDOLIZUMAB

Code 86711

Lot No. 860



2023-08



2022-08

REAGENTS

Label	Description	Lot.	Exp.	NO
D4490	PROMONITOR ANTI-VEDOLIZUMAB	860	2023-08	6x6
B4490	CALIBRATORE ANTI-VEDOLIZUMAB	860	2024-06	1
C4490	CONTROL SERUM ANTI-VEDOLIZUMAB	860	2024-06	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	330	≥ 100
Control Serum (AU/mL)	99,6	68,5-114,1

CALIBRATOR


2 7 9 8 6 0 6 4 1 9 3 1 3 2 5 1 0 0

Label checked with barcode reader

YES



Operator Signature *AB*



CONTROL SERUM


2 7 9 8 6 0 6 2 5 3 8 0 0 0 0 0 0 1

Label checked with barcode reader

YES



Operator Signature *AB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Alc...*

DATE *05-12-2021*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.