

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: 88000-F95 PEACH

Code	88000-F95		
Lot No.	878	2023-10	2022-12

REAGENTS

Label	Description	Lot	Exp	No
D5751	88000-F95 PEACH	878	2023-10	2X6
B5750	CALIBRATORE 88000-F95 PEACH	878	2024-02	1
C5750	CONTROL SERUM 88000-F95 PEACH	878	2024-02	1


CONTROLS

	Found results	Expected result
Calibrator (mOD)	256	≥ 100
Control Serum (KU/L)	13,1	7,9-17,6

CALIBRATOR

6	4	7	8	7	8	6	5	1	8	5	5	2	2	9	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *SB*



CONTROL SERUM

6	4	7	8	7	8	6	2	9	5	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *SB*



The test has been carried out in accordance with the package insert

TESTED BY: Simone Baiocchi *Simone Baiocchi*
APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*
DATE *13.12.2022*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.